

HiMedia Laboratories Private Limited

C-40, Road No.21Y, MIDC, Wagle Industrial Area,
Thane(W) - 400604 , Website : www.himedialabs.com,
Email : info@himedialabs.com

Certificate of Analysis, Quality and Conformity

Material Code : M1418	Material Name : *HiCrome™ UTI Agar, Modified	Lot No : 000570140
Report No.: 40001315014	Date of Release & Report : 2023-01-04	Expiry Date : 2025-12

Appearance

Cream to yellow homogeneous free flowing powder. Observed : Light yellow

Gelling

Firm, comparable with 1.5% Agar gel

Colour and Clarity of prepared medium

Light amber coloured, clear to slightly opalescent gel forms in Petri plates

Reaction

Reaction of 5.54% w/v aqueous solution at 25°C.

pH

pH Range :7.00-7.40 Observed : 7.38

Cultural Response

Cultural characteristics observed after an incubation at 35-37°C for 24 hours.

Organism	Inoculum (CFU)	Growth	Recovery	Colour of Colony	TDA (add 1-2 drops of TDA reagent)	DMACA (transfer colony on filter paper d
Cultural Response						
<i>Escherichia coli</i> ATCC 25922 (WDCM 00013)	50-100	luxuriant	>=70%	Purple to magenta	negative reaction	positive reaction, formation of blue purple colour around growth
<i>Enterococcus faecalis</i> ATCC 29212 (WDCM 00087)	50-100	luxuriant	>=70%	blue-green (small)	negative reaction	negative reaction
<i>Klebsiella pneumoniae</i> ATCC 13883 (WDCM 00097)	50-100	luxuriant	>=70%	blue to purple, mucoid	negative reaction	negative reaction
<i>Proteus mirabilis</i> ATCC 12453	50-100	luxuriant	>=70%	light brown	positive reaction, development of brown colouration	negative reaction
<i>Pseudomonas aeruginosa</i> ATCC 27853 (WDCM 00025)	50-100	luxuriant	>=70%	colourless (greenish pigment may be observed)	negative reaction	negative reaction
<i>Staphylococcus aureus</i> ATCC 25923 (WDCM 00034)	50-100	luxuriant	>=70%	golden yellow	negative reaction	negative reaction

- . ATCC is a registered trade mark of the American Type Culture Collection
- . NCTC and National Collection of Type Culture are registered trade mark of the Health Protection Agency

HiMedia Laboratories Private Limited

C-40, Road No.21Y, MIDC, Wagle Industrial Area,
Thane(W) - 400604 , Website : www.himedialabs.com,
Email : info@himedialabs.com

Certificate of Analysis, Quality and Conformity

Material Code : M1418	Material Name : *HiCrome™ UTI Agar, Modified	Lot No : 000570140
Report No.: 40001315014	Date of Release & Report : 2023-01-04	Expiry Date : 2025-12

Control Media :

- . For Bacteria : Soyabean Casein Digest Agar / Columbia Blood Agar base enriched with 5% v/v Sheep/Horse blood.
- . For Yeast & Mold : Sabouraud Dextrose Agar.

- . All ISO 11133 : 2014/Amd.1:2018(E) control strains are included in the Quality parameter
- . HiMedia Laboratories Pvt Ltd is Certified for ISO 9001:2015, ISO 13485:2016 , WHO GMP

. Information for BSE/TSE Risk: The material was subjected to pH <= 7.0 and/or a temperature in excess of 75°C for no less than 2 hours during the manufacturing process. The bovine raw material for this product was collected entirely from Indian Origin animals in a licensed based establishment. The animals are inspected under a Govt. approved veterinarian's supervision and were apparently free from infectious and contagious diseases. BSE (Bovine Spongiform Encephalopathy)/ TSE (Transmissible Spongiform Encephalopathy) and dioxine are not known to exist in India. This material does not contain, nor is derived from the specific risks material as defined in The Maharashtra Animal Preservation Act Govt. of Maharashtra, India.

STATUS OF THE MATERIAL : APPROVED

This is to certify that this lot passes and it confirms to the above mentioned tests and specifications . The information given here is believed to be correct and accurate, however, both the information and products are offered without warranty for any particulars use, other than that specified in the current HiMedia manual or product sheets. The results reported were obtained at the time of release.

This document has been produced electronically and is valid


Maya Sonavane

**Microbiologist/Sr.Executive
Microbiologist**


Ujwala M. Kokate

Asst./Dy/QC Manager


Dr. Santosh Kaul

Dy/QA Manager

2023-01-04