

HiMedia Laboratories Private Limited

23, Vadhani Industrial Estate, L.B.S. Marg,
Mumbai - 400086 , Website : www.himedialabs.com,
Email : info@himedialabs.com

Certificate of Analysis, Quality and Conformity

Material Code : M053	Material Name : Phenol Red Agar Base	Lot No : 0000428785
Report No.: 40001014472	Date of Release & Report : 2020-04-22	Expiry Date : 2025-02

Appearance

Light yellow to pink homogeneous Free flowing powder . Observed : Light pink

Gelling

firm, comparable with 1.5% Agar gel

Colour and Clarity of prepared medium

Red coloured clear to slightly opalescent gel forms in tubes as slants

Reaction

Reaction of 3.1% w/v aqueous solution at 25°C.

pH

pH Range :7.20-7.60 Observed : 7.46

Cultural Response

Cultural characteristics observed after an incubation at 35 - 37°C for 18 - 24 hours.

Organism	Inoculum (CFU)	Growth	without carbohydrate, (Acid)	without carbohydrate, (Gas)	with dextrose,(Acid)	with dextrose,(Gas)
Cultural Response						
<i>Alcaligenes faecalis</i> ATCC 8750	50-100	luxuriant	Negative reaction, no colour change	Negative reaction	Negative reaction, no colour change	Negative reaction
<i>Escherichia coli</i> ATCC 25922 (WDCM 00013)	50-100	luxuriant	Negative reaction, no colour change	Negative reaction	Positive reaction, yellow colour	Positive reaction
<i>Klebsiella pneumoniae</i> ATCC 13883 (WDCM 00097)	50-100	luxuriant	Negative reaction, no colour change	Negative reaction	Positive reaction, yellow colour	Positive reaction
<i>Proteus hauseri</i> ATCC 13315	50-100	luxuriant	Negative reaction, no colour change	Negative reaction	Positive reaction, yellow colour	Positive reaction
<i>Salmonella enterica</i> serovar Typhimurium ATCC 14028 (WDCM 00031)	50-100	luxuriant	Negative reaction, no colour change	Negative reaction	Positive reaction, yellow colour	Positive reaction
<i>Shigella flexneri</i> ATCC 12022 (WDCM 00126)	50-100	luxuriant	Negative reaction, no colour change	Negative reaction	Positive reaction, yellow colour	Negative reaction

- . ATCC is a registered trade mark of the American Type Culture Collection
- . NCTC and National Collection of Type Culture are registered trade mark of the Health Protection Agency

HiMedia Laboratories Private Limited

23, Vadhani Industrial Estate, L.B.S. Marg,
Mumbai - 400086 , Website : www.himedialabs.com,
Email : info@himedialabs.com

Certificate of Analysis, Quality and Conformity

Material Code : M053	Material Name : Phenol Red Agar Base	Lot No : 0000428785
Report No.: 40001014472	Date of Release & Report : 2020-04-22	Expiry Date : 2025-02

Control Media :

- . For Bacteria : Soyabean Casein Digest Agar / Columbia Blood Agar base enriched with 5% v/v Sheep/Horse blood.
- . For Yeast & Mold : Sabouraud Dextrose Agar.


- . All ISO 11133 : 2014/Amd.1:2018(E) control strains are included in the Quality parameter
- . HiMedia Laboratories Pvt Ltd is Certified for ISO 9001:2015, ISO 13485:2016 , WHO GMP

. Information for BSE/TSE Risk: The material was subjected to pH \leq 7.0 and/or a temperature in excess of 75°C for no less than 2 hours during the manufacturing process. The bovine raw material for this product was collected entirely from Indian Origin animals in a licensed based establishment. The animals are inspected under a Govt. approved veterinarian's supervision and were apparently free from infectious and contagious diseases. BSE (Bovine Spongiform Encephalopathy)/ TSE (Transmissible Spongiform Encephalopathy) and dioxine are not known to exist in India. This material does not contain, nor is derived from the specific risks material as defined in The Maharashtra Animal Preservation Act Govt. of Maharashtra, India.

STATUS OF THE MATERIAL : APPROVED

This is to certify that this lot passes and it confirms to the above mentioned tests and specifications . The information given here is believed to be correct and accurate, however, both the information and products are offered without warranty for any particulars use, other than that specified in the current HiMedia manual or product sheets. The results reported were obtained at the time of release.

This document has been produced electronically and is valid



Sheetal Shewale

**Microbiologist/Sr.Executive
Microbiologist**



Ujwala M. Kokate

Asst./Dy/QC Manager



Dr. Santosh Kaul

Dy/QA Manager

2020-04-22