



Listeria Identification Agar Base (PALCAM)

M1064

Intended use

Recommended for selective isolation and identification of *Listeria* species from clinical and non-clinical samples.

Composition**

Ingredients	g / L
Peptone	23.000
Starch	1.000
Sodium chloride	5.000
Mannitol	10.000
Ammonium ferric citrate	0.500
Esculin	0.800
Dextrose (Glucose)	0.500
Lithium chloride	15.000
Phenol red	0.080
Agar	13.000
Final pH (at 25°C)	7.0±0.2

**Formula adjusted, standardized to suit performance parameters

Directions

Suspend 34.44 grams in 500 ml purified/distilled water. Heat to boiling to dissolve the medium completely. Sterilize by autoclaving at 15 lbs pressure (121°C) for 15 minutes. Cool to 45-50°C and aseptically add rehydrated contents of 1 vial of PALCAM Selective Supplement (FD061). Mix well and pour into sterile Petri plates.

Principle And Interpretation

The genus *Listeria* constitutes *Listeria monocytogenes*, *Listeria ivanovii*, *Listeria seeligeri*, *Listeria welshimerii*, *Listeria innocua*, *Listeria grayi*, *Listeria murrayi* and *Listeria denitrificans*. Among these, *L. monocytogenes* and *L. ivanovii* are associated with diseases in humans. The pathogenicity of *L. ivanovii* is uncertain. *L. monocytogenes* is found in a wide variety of habitats, including the normal microflora of healthy ruminants, gastrointestinal tract of asymptomatic humans and environmental sources including river water, sewage, soil, silage, fertilizers and decaying vegetation (1).

Listeria Identification Agar also known as Polymyxin-Acriflavin-Lithium chloride-Ceftazidime-Aesculin-Mannitol (PALCAM) Agar was formulated by Van Netten et al (2) and is recommended for the isolation of *L. monocytogenes* from foods. PALCAM medium is highly selective due to the presence of lithium chloride, ceftazidime, polymyxin B and acriflavin hydrochloride. PALCAM medium is a differential diagnostic medium utilizing two indicator systems, as esculin and ferric citrate and mannitol and phenol red.

Peptone serves as carbon, nitrogen substances, long chain amino acids, vitamins and essential growth nutrients for the organisms. Dextrose (Glucose), starch and mannitol are the carbohydrate and energy sources. Sodium chloride maintains the osmotic equilibrium of the medium. Phenol red is the pH indicator dye that exhibits changes in the pH of the medium. *L. monocytogenes* hydrolyzes esculin to form esculetin and dextrose. Esculetin reacts with ammonium ferric citrate and forms a brown-black complex seen as a black halo around colonies. *L. monocytogenes* does not ferment mannitol but contaminants such as Enterococci and Staphylococci ferment mannitol and is indicated by colour change from red to yellow. Under microaerophilic conditions, strict aerobes such as *Bacillus* species and *Pseudomonas* species are inhibited. The addition of egg yolk (2.5% v/v) to PALCAM Agar has been reported to aid repair of damaged cells (3). Medium containing blood when overlaid on PALCAM Agar enables to differentiate and enumerate haemolytic *Listeria* species (4). Depending upon the type of sample used, selective enrichment broth should be used prior to inoculation onto PALCAM Agar. Generally *Listeria* Selective Enrichment Medium is used for dairy products and *Listeria* Selective Enrichment Medium UVM (M890A), Fraser Secondary Enrichment Broth (M1083) are used for meats and poultry. On PALCAM Agar, colonies of *Listeria* appear as grey-green with a black precipitate, following inoculation and incubation at 35°C for 24-48 hours under aerobic or microaerophilic conditions.

Type of specimen

Clinical samples - faeces , body fluids, Food samples; Water samples

Specimen Collection and Handling:

For clinical samples follow appropriate techniques for handling specimens as per established guidelines (5,6).

For food samples, follow appropriate techniques for sample collection and processing as per guidelines (7).

For water samples, follow appropriate techniques for sample collection, processing as per guidelines and local standards (8).

After use, contaminated materials must be sterilized by autoclaving before discarding.

Warning and Precautions

In Vitro diagnostic Use. For professional use only. Read the label before opening the container. Wear protective gloves/protective clothing/eye protection/face protection. Follow good microbiological lab practices while handling specimens and culture. Standard precautions as per established guidelines should be followed while handling clinical specimens. Safety guidelines may be referred in individual safety data sheets.

Limitations :

1. Individual organisms differ in their growth requirement and may show variable growth patterns on the medium.
2. Each lot of the medium has been tested for the organisms specified on the COA. It is recommended to users to validate the medium for any specific microorganism other than mentioned in the COA based on the user's unique requirement.
3. The medium is not differential, so further biochemical testing is required for identification between *Listeria* species.

Performance and Evaluation

Performance of the medium is expected when used as per the direction on the label within the expiry period when stored at recommended temperature.

Quality Control

Appearance

Light yellow to pink homogeneous free flowing powder

Gelling

Firm, comparable with 1.3% Agar gel.

Colour and Clarity of prepared medium

Red coloured clear to slightly opalescent gel forms in Petri plates.

Reaction

Reaction of 6.9% w/v aqueous solution at 25°C. pH : 7.0±0.2

pH

6.80-7.20

Cultural Response

Cultural characteristics observed under microaerophilic condition, with added PALCAM Selective Supplement (FD061), after an incubation at 35-37°C for 24-48 hours.

Organism	Inoculum (CFU)	Growth	Recovery	Colony characteristics
<i>Enterococcus faecalis</i> ATCC 29212 (00087*)	50-100	none-poor	≤10%	grey colonies with a brown-green halo
<i>Listeria monocytogenes</i> ATCC 19111 (00020*)	50-100	luxuriant	≥50%	grey-green with black center and a black halo
<i>Listeria monocytogenes</i> ATCC 19112	50-100	luxuriant	≥50%	grey-green with black center and a black halo
<i>Listeria monocytogenes</i> ATCC 19117	50-100	luxuriant	≥50%	grey-green with black center and a black halo
<i>Listeria monocytogenes</i> ATCC 19118	50-100	luxuriant	≥50%	grey-green with black center and a black halo

<i>Staphylococcus aureus</i> <i>subsp. aureus</i> ATCC 25923 (00034*)	50-100	none-poor	<=10%	yellow colonies with yellow halo
---	--------	-----------	-------	-------------------------------------

Key : (*) Corresponding WDCM numbers.

Storage and Shelf Life

Store between 10-30°C in a tightly closed container and the prepared medium at 2-8°C. Use before expiry date on the label. On opening, product should be properly stored dry, after tightly capping the bottle in order to prevent lump formation due to the hygroscopic nature of the product. Improper storage of the product may lead to lump formation. Store in dry ventilated area protected from extremes of temperature and sources of ignition. Seal the container tightly after use. Product performance is best if used within stated expiry period.

Disposal

User must ensure safe disposal by autoclaving and/or incineration of used or unusable preparations of this product. Follow established laboratory procedures in disposing of infectious materials and material that comes into contact with clinical sample must be decontaminated and disposed of in accordance with current laboratory techniques (5,6).

Reference

1. Watkin J., Sleath K. P., J. Appl. Bacteriol., 50: 1-9, 1981.
2. Van Netten P., Peralse I, Van de Mosdik A., Curtis G.D.W., Mossel D. A.A., 1989, Int. J. Food Microbiol., 8(4):299.
3. Veld P.H. and de Boer E., 1991, Int. J. Food Microbiol., 13:295.
4. Van Netten P., van Gaal B. and Mossel D. A. A., 1991, Lett. Appl.Microbiol, 12:20.
5. Isenberg, H.D. Clinical Microbiology Procedures Handbook 2nd Edition.
6. Jorgensen, J.H., Pfaller, M.A., Carroll, K.C., Funke, G., Landry, M.L., Richter, S.S and Warnock., D.W. (2015) Manual of Clinical Microbiology, 11th Edition. Vol. 1.
7. Salfinger Y., and Tortorello M.L. Fifth (Ed.), 2015, Compendium of Methods for the Microbiological Examination of Foods, American Public Health Association, Washington, D.C.
8. Lipps WC, Braun-Howland EB, Baxter TE, eds. Standard methods for the Examination of Water and Wastewater, 24th ed. Washington DC:APHA Press; 2023.


Revision : 05/2024



HiMedia Laboratories Pvt. Limited,
Plot No.C-40, Road No.21Y,
MIDC, Wagle Industrial Area,
Thane (W) -400604, MS, India



**In vitro diagnostic
medical device**



Storage temperature



CEpartner4U, Esdoornlaan 13,
3951DB Maarn, NL
www.cepartner4u.eu



CE Marking



**Do not use if
package is damaged**

Disclaimer :

User must ensure suitability of the product(s) in their application prior to use. Products conform solely to the information contained in this and other related HiMedia™ publications. The information contained in this publication is based on our research and development work and is to the best of our knowledge true and accurate. HiMedia™ Laboratories Pvt Ltd reserves the right to make changes to specifications and information related to the products at any time. Products are not intended for human or animal or therapeutic use but for laboratory, diagnostic, research or further manufacturing use only, unless otherwise specified. Statements contained herein should not be considered as a warranty of any kind, expressed or implied, and no liability is accepted for infringement of any patents.