

HiMedia Laboratories Private Limited

23, Vadhani Industrial Estate, L.B.S. Marg,
Mumbai - 400086 , Website : www.himedialabs.com,
Email : info@himedialabs.com

Certificate of Analysis, Quality and Conformity

Material Code : M1075	Material Name : Endo Agar, Modified	Lot No : 0000489223
Report No.: 40001138971	Date of Release & Report : 2021-07-07	Expiry Date : 2026-06

Appearance

Light pink to purple homogeneous free flowing powder . Observed : Light purple

Gelling

Firm, comparable with 1.25% Agar gel

Colour and Clarity of prepared medium

Orangish pink coloured, clear to slightly opalescent gel with fine precipitate forms in Petri plates.

Reaction

Reaction of 3.86% w/v aqueous solution at 25°C.

pH

pH Range :7.20-7.60 Observed : 7.52

Cultural Response

Cultural characteristics observed after an incubation at 35-37°C for 18- 24 hours.

Organism	Inoculum (CFU)	Growth	Recovery	Colour of Colony
Cultural Response				
<i>Bacillus spizizenii</i> ATCC 6633 (WDCM 00003)	>=10 ⁴	inhibited	0%	-
<i>Enterobacter aerogenes</i> ATCC 13048 (WDCM 00175)	50-100	good-luxuriant	>=50%	pink
<i>Enterococcus faecalis</i> ATCC 29212 (WDCM 00087)	50-100	none-poor	<=10%	pink, small
<i>Escherichia coli</i> ATCC 25922 (WDCM 00013)	50-100	good-luxuriant	>=50%	pink to rose red with metallic sheen
<i>Klebsiella pneumoniae</i> ATCC 13883 (WDCM 00097)	50-100	good-luxuriant	>=50%	pink, mucoid
<i>Proteus hauseri</i> ATCC 13315	50-100	good-luxuriant	>=50%	colourless to pale pink
<i>Pseudomonas aeruginosa</i> ATCC 27853 (WDCM 00025)	50-100	good-luxuriant	>=50%	colourless, irregular
<i>Salmonella enterica</i> serovar Typhi ATCC 6539	50-100	good-luxuriant	>=50%	colourless to pale pink
<i>Shigella sonnei</i> ATCC 25931	50-100	good-luxuriant	>=50%	colourless to pale pink
<i>Staphylococcus aureus</i> ATCC 25923 (WDCM 00034)	>=10 ⁴	inhibited	0%	-
<i>Enterobacter cloacae</i> ATCC	50-100	good	40-50%	pink

HiMedia Laboratories Private Limited

23, Vadhani Industrial Estate, L.B.S. Marg,
Mumbai - 400086 , Website : www.himedialabs.com,
Email : info@himedialabs.com

Certificate of Analysis, Quality and Conformity

Material Code : M1075	Material Name : Endo Agar, Modified	Lot No : 0000489223
Report No.: 40001138971	Date of Release & Report : 2021-07-07	Expiry Date : 2026-06

<i>13047 (WDCM 00083)</i>				
<i>Salmonella enterica serovar Typhimurium ATCC 14028 (WDCM 00031)</i>	50-100	good-luxuriant	>=50%	colourless
<i>Salmonella enterica serovar Enteritidis ATCC 13076 (WDCM 00030)</i>	50-100	good-luxuriant	>=50%	colourless
<i>Shigella flexneri ATCC 12022 (WDCM 00126)</i>	50-100	good-luxuriant	>=50%	colourless

- . ATCC is a registered trade mark of the American Type Culture Collection
- . NCTC and National Collection of Type Culture are registered trade mark of the Health Protection Agency

Control Media :

- . For Bacteria : Soyabean Casein Digest Agar / Columbia Blood Agar base enriched with 5% v/v Sheep/Horse blood.
- . For Yeast & Mold : Sabouraud Dextrose Agar.
- . All ISO 11133 : 2014/Amd.1:2018(E) control strains are included in the Quality parameter
- . HiMedia Laboratories Pvt Ltd is Certified for ISO 9001:2015, ISO 13485:2016 , WHO GMP

. Information for BSE/TSE Risk: The material was subjected to pH <= 7.0 and/or a temperature in excess of 75°C for no less than 2 hours during the manufacturing process. The bovine raw material for this product was collected entirely from Indian Origin animals in a licensed based establishment. The animals are inspected under a Govt. approved veterinarian's supervision and were apparently free from infectious and contagious diseases. BSE (Bovine Spongiform Encephalopathy)/ TSE (Transmissible Spongiform Encephalopathy) and dioxine are not known to exist in India. This material does not contain, nor is derived from the specific risks material as defined in The Maharashtra Animal Preservation Act Govt. of Maharashtra, India.

STATUS OF THE MATERIAL : APPROVED

This is to certify that this lot passes and it confirms to the above mentioned tests and specifications . The information given here is believed to be correct and accurate, however, both the information and products are offered without warranty for any particulars use, other than that specified in the current HiMedia manual or product sheets. The results reported were obtained at the time of release.

This document has been produced electronically and is valid



Maya Sonavane

Microbiologist/Sr.Executive
Microbiologist



Ujwala M. Kokate

Asst./Dy/QC Manager



Dr. Santosh Kaul

Dy/QA Manager

2021-07-07