

**HiMedia Laboratories Private Limited**

C-40, Road No.21Y, MIDC, Wagle Industrial Area,  
Thane(W) - 400604 , Website : www.himedialabs.com,  
Email : info@himedialabs.com

**Certificate of Analysis, Quality and Conformity**

<b>Material Code : M1417F</b>	<b>Material Name :</b> HiCrome™ Listeria Agar Base	<b>Lot No : 000583976</b>
<b>Report No.: 40001346067</b>	<b>Date of Release &amp; Report : 2023-04-12</b>	<b>Expiry Date : 2026-03</b>

**Appearance**

Light yellow to pink homogeneous free flowing powder. Observed : Pinkish beige

**Gelling**

Firm, comparable with 1.3% Agar gel.

**Colour and Clarity of prepared medium**

Red coloured, clear to slightly opalescent gel forms in Petri plates

**Reaction**

Reaction of 7.33% w/v aqueous solution at 25°C.

**pH**

pH Range :7.20-7.40 Observed : 7.38

**Cultural Response**

Cultural characteristics observed w/added HiCrome Listeria Selective Supplement (FD181), after an incubation at 35-37°C for 24-48 hours.

Organism	Inoculum (CFU)	Growth	Recovery	Colour of colony	Rhamnose fermentation
<b>Cultural Response</b>					
<i>Bacillus spizizenii</i> ATCC 6633 (WDCM 00003)	>=10 <sup>4</sup>	inhibited	0%-	-	-
<i>Candida albicans</i> ATCC 10231 (WDCM 00054)	>=10 <sup>4</sup>	inhibited	0%	-	-
<i>Escherichia coli</i> ATCC 25922 (WDCM 00013)	>=10 <sup>4</sup>	inhibited	0%	-	-
<i>Listeria innocua</i> ATCC 33090 (WDCM 00017)	50-100	luxuriant	>=50%	bluish green	positive reaction, (yellow background)
<i>Listeria ivanovii</i> ATCC 19119 (WDCM 00018)	50-100	luxuriant	>=50%	bluish green	negative reaction
<i>Listeria monocytogenes</i> ATCC 19118	50-100	luxuriant	>=50%	bluish green	positive reaction,(yellow halo)
<i>Pseudomonas aeruginosa</i> ATCC 27853 (WDCM 00025)	>=10 <sup>4</sup>	inhibited	0%	-	-

- . ATCC is a registered trade mark of the American Type Culture Collection
- . NCTC and National Collection of Type Culture are registered trade mark of the Health Protection Agency

**Control Media :**

- . For Bacteria : Soyabean Casein Digest Agar / Columbia Blood Agar base enriched with 5% v/v Sheep/Horse blood.

**HiMedia Laboratories Private Limited**

C-40, Road No.21Y, MIDC, Wagle Industrial Area,  
Thane(W) - 400604 , Website : www.himedialabs.com,  
Email : info@himedialabs.com

**Certificate of Analysis, Quality and Conformity**

<b>Material Code : M1417F</b>	<b>Material Name :</b> HiCrome™ Listeria Agar Base	<b>Lot No : 000583976</b>
<b>Report No.: 40001346067</b>	<b>Date of Release &amp; Report : 2023-04-12</b>	<b>Expiry Date : 2026-03</b>

. For Yeast & Mold : Sabouraud Dextrose Agar.


. All ISO 11133 : 2014/Amd.1:2018( E ) control strains are included in the Quality parameter  
. HiMedia Laboratories Pvt Ltd is Certified for ISO 9001:2015, ISO 13485:2016 , WHO GMP

. Information for BSE/TSE Risk: The material was subjected to pH  $\leq$  7.0 and/or a temperature in excess of 75°C for no less than 2 hours during the manufacturing process. The bovine raw material for this product was collected entirely from Indian Origin animals in a licensed based establishment. The animals are inspected under a Govt. approved veterinarian's supervision and were apparently free from infectious and contagious diseases. BSE (Bovine Spongiform Encephalopathy)/ TSE (Transmissible Spongiform Encephalopathy) and dioxine are not known to exist in India. This material does not contain, nor is derived from the specific risks material as defined in The Maharashtra Animal Preservation Act Govt. of Maharashtra, India.

**STATUS OF THE MATERIAL : APPROVED**

This is to certify that this lot passes and it confirms to the above mentioned tests and specifications . The information given here is believed to be correct and accurate, however, both the information and products are offered without warranty for any particulars use, other than that specified in the current HiMedia manual or product sheets. The results reported were obtained at the time of release.

**This document has been produced electronically and is valid**



Siddhi Alegaonkar

**Microbiologist/Sr.Executive  
Microbiologist**



Ujwala M. Kokate

**Asst./Dy/QC Manager**



Dr. Santosh Kaul

**Dy/QA Manager**

2023-04-12